



SCIVISION
BIOTECH

SciVision Biotech Inc.

9, South 6th Rd., K.E.P.Z., Kaohsiung 806, Taiwan, R.O.C.

<http://www.scivisionbiotech.com>

E-mail: service@scivision.com.tw

Tel: +886-7-823-2258

Fax: +886-7-823-2295

CERTIFICATE OF ANALYSIS

I, responsible QA head in plant, SciVision, certify that the enclosed are the results of analysis for the product specified below:

Product : **Flexotron FORTE Synovial Fluid Supplement**

Pre filled syringes, each 3.0 ml containing 30 mg Sodium Hyaluronate

Lot No. : **MDFE-AYAB**

Manufacturing Date : 28-May-2025

Expiry Date : 27-May-2028

Test Item	Specification	Result
Description	Clear, colorless, viscous liquid filled in pre-filled syringe	PASS
pH	6.8 ~ 7.8	7.5
Osmolality	270 ~ 390 mOsm/kg	336 mOsm/kg
Bacterial Endotoxins	less than 0.25 EU/mL	PASS
Sterility	Sterile	Sterile
Assay	0.9 ~ 1.2 %	1.0 %
Active Volume	2.8 ~ 3.2 g	PASS
Intrinsic Viscosity	1.305~ 1.990 m ³ /kg	1.898 m ³ /kg
Extrusion Force	5~30 N	PASS

Signature

Patty Tsai

Name

Patty Tsai

Date Issued

2025/07/14





Документ подписан и передан через оператора ЭДО АО «ПФ «СКБ Контур»

	Организация, сотрудник	Доверенность: рег. номер, период действия и статус	Сертификат: серийный номер, период действия	Дата и время подписания
Подписи отправителя:	 Общество с ограниченной ответственностью "МКНТ Импорт" Егорова Ирина Викторовна Доверитель: ОБЩЕСТВО С ОГРАНИЧЕННОЙ ОТВЕТСТВЕННОСТЬЮ "МКНТ ИМПОРТ"	 3cc4bde2-d025-4bbe-b04b-e5e61df94798 с 07.04.2025 00:00 по 07.04.2026 23:59 GMT+03:00 Доверенность прошла проверку	06DD048100C6B2BFA94F692F2 FA97DBCA0 с 21.04.2025 10:44 по 21.04.2026 10:49 GMT+03:00	16.09.2025 16:00 GMT+03:00 Подпись соответствует файлу документа